

EC Certificate

Directive 98/79/EC Annex IV, excluding Sections 4 and 6
Full Quality Assurance System
In Vitro Diagnostic Medical Devices

Registration No.: HL 60105035 0001

Report No.: 21235785 002

Manufacturer: Van Oostveen Medical BV
Herenweg 269
3648 CH Wilnis
Netherlands

Products: In-vitro Diagnostic Tests for self-testing for
Blood Glucose Monitoring and Pregnancy

Expiry Date: 2020-10-07

The Notified Body hereby declares that the requirements of Annex IV, excluding section 4 and 6 of the directive 98/79/EC have been met for the listed products. The above named manufacturer has established and applies a quality assurance system, which is subject to periodic surveillance, defined by Annex IV, section 5 of the aforementioned directive. For placing on the market of List A devices covered by this certificate an EC design-examination certificate according to Annex IV, section 4 and a verification of manufactured products according to section 6 is required.

Effective Date: 2016-04-11

Date: 2016-04-11



TÜV Rheinland LGA Products GmbH - Tillystraße 2 - 90431 Nürnberg

TÜV Rheinland LGA Products GmbH is a Notified Body according to Directive 98/79/EC concerning in vitro diagnostic medical devices with the identification number 0197.